

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 12/02/2019

OPERATOR NAME:
DILLAN L CHERY
Permit #: 42470
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:24
INTERNAL STANDARD	VERIFIED	15:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:25
BLANK TEST	0.00	15:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27

