

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 12/09/2019

OPERATOR NAME:
KHALID PARWANTA
Permit #: 31498
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:40
INTERNAL STANDARD	VERIFIED	17:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:41
BLANK TEST	0.00	17:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44

