

SUPERVISOR TEST

-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 12/16/2019

OPERATOR NAME:  
KHALID PARWANTA  
Permit #: 31498  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:29
INTERNAL STANDARD	VERIFIED	12:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:31
BLANK TEST	0.00	12:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:32
BLANK TEST	0.00	12:33

