

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 12/30/2019

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:29
INTERNAL STANDARD	VERIFIED	10:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:29
BLANK TEST	0.00	10:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:30
BLANK TEST	0.00	10:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:31
BLANK TEST	0.00	10:32

P.O.
[Signature]
[Signature]

