

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 01/06/2020

OPERATOR NAME:
DILLAN L CHERY
Permit #: 42470
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:37
INTERNAL STANDARD	VERIFIED	20:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:38
BLANK TEST	0.00	20:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:39
BLANK TEST	0.00	20:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41

