

SUPERVISOR TEST

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 01/13/2020

OPERATOR NAME:
 DILLAN L CHERY
 Permit #: 42470
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:30
INTERNAL STANDARD	VERIFIED	13:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:30
BLANK TEST	0.00	13:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:31
BLANK TEST	0.00	13:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:33

