

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 01/20/2020

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:45
INTERNAL STANDARD	VERIFIED	22:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:45
BLANK TEST	0.00	22:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:48
BLANK TEST	0.00	22:49

