

SUPERVISOR TEST

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 02/24/2020

OPERATOR NAME:
 PHILIP E MISSERE
 Permit #: 41131
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:36
INTERNAL STANDARD	VERIFIED	10:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:36
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:38
BLANK TEST	0.00	10:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:40

