

SOLUTION CHANGE

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 02/26/2020

OPERATOR NAME:  
BRANDON KING  
Permit #: 36388  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:55
INTERNAL STANDARD	VERIFIED	11:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:56
BLANK TEST	0.00	11:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:57
BLANK TEST	0.00	11:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:58
BLANK TEST	0.00	11:59

