

SUPERVISOR TEST

-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 03/09/2020

OPERATOR NAME:  
MATTHEW J BOWE  
Permit #: 40643  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:07
INTERNAL STANDARD	VERIFIED	09:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:10
BLANK TEST	0.00	09:11

P.O.  
*[Signature]* #620

