

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 03/23/2020

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:07
INTERNAL STANDARD	VERIFIED	09:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:07
BLANK TEST	0.00	09:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:10
BLANK TEST	0.00	09:11

MJB
#620

