

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 04/06/2020

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:53
INTERNAL STANDARD	VERIFIED	08:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:54
BLANK TEST	0.00	08:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:56
BLANK TEST	0.00	08:57

M J Bowe
#040

