

SUPERVISOR TEST

-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 04/13/2020

OPERATOR NAME:  
SIMONE FONVIL  
Permit #: 36335  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:01
INTERNAL STANDARD	VERIFIED	11:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:02
BLANK TEST	0.00	11:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:03
BLANK TEST	0.00	11:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:04
BLANK TEST	0.00	11:05

