

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 05/05/2020

OPERATOR NAME:
KHALID PARWANTA
Permit #: 31498
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:16
INTERNAL STANDARD	VERIFIED	15:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:16
BLANK TEST	0.00	15:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:18
BLANK TEST	0.00	15:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:19
BLANK TEST	0.00	15:20

