

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 05/11/2020

OPERATOR NAME:
KHALID PARWANTA
Permit #: 31498
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:32
INTERNAL STANDARD	VERIFIED	10:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:32
BLANK TEST	0.00	10:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36

