

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 05/18/2020

OPERATOR NAME:
PHILIP MISSERE
Permit #: 41131
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:44 |
| INTERNAL STANDARD | VERIFIED | 08:45 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:45 |
| BLANK TEST | 0.00 | 08:46 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:46 |
| BLANK TEST | 0.00 | 08:47 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:47 |
| BLANK TEST | 0.00 | 08:48 |

