

SUPERVISOR TEST

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 06/29/2020

OPERATOR NAME:
 ANDRE R RIOBE
 Permit #: 47395
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:00
INTERNAL STANDARD	VERIFIED	14:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:03
BLANK TEST	0.00	14:04

