

SOLUTION CHANGE




SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 07/07/2020

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:03
BLANK TEST	0.00	09:04



[Handwritten signature]
[Handwritten initials]