

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 07/13/2020

OPERATOR NAME:
ANDREW S COLE-HATCHARD
Permit #: 47565
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:22
INTERNAL STANDARD	VERIFIED	13:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:22
BLANK TEST	0.00	13:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:24
BLANK TEST	0.00	13:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:25
BLANK TEST	0.00	13:26

