

SUPERVISOR TEST

-----  
 SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106  
 -----

Date: 07/20/2020

OPERATOR NAME:  
 MATTHEW J BOWE  
 Permit #: 40643  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:38
BLANK TEST	0.00	08:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:39
BLANK TEST	0.00	08:40

*Bo*  
*MJE*  
*#Cadd*

