

SUPERVISOR TEST

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 07/27/2020

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:11
INTERNAL STANDARD	VERIFIED	08:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:12
BLANK TEST	0.00	08:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15

