

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 08/10/2020

OPERATOR NAME:
 PHILIP MISSERE
 Permit #: 41131
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:52
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:54
BLANK TEST	0.00	11:55

