

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 08/17/2020

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:01
INTERNAL STANDARD	VERIFIED	09:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:04
BLANK TEST	0.00	09:05

P.O. [Signature]
 #020

