

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 08/24/2020

OPERATOR NAME:
 MATHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:11
BLANK TEST	0.00	17:12

Handwritten signatures:
 JCB
 HCB

