

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 08/31/2020

OPERATOR NAME:
 BANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:51
INTERNAL STANDARD	VERIFIED	17:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:51
BLANK TEST	0.00	17:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:53
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:54
BLANK TEST	0.00	17:55

