

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 09/07/2020

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31

