

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 09/21/2020

OPERATOR NAME:  
DILLAN L CHERY  
Permit #: 42470  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:27
INTERNAL STANDARD	VERIFIED	11:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:28
BLANK TEST	0.00	11:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:30
BLANK TEST	0.00	11:31

