

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 09/29/2020

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:45
INTERNAL STANDARD	VERIFIED	21:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:46
BLANK TEST	0.00	21:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:47
BLANK TEST	0.00	21:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:48
BLANK TEST	0.00	21:49

