

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106

Date: 10/05/2020

OPERATOR NAME:  
 MATTHEW J BOWE  
 Permit #: 40643  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:35
INTERNAL STANDARD	VERIFIED	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:36
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:37
BLANK TEST	0.00	10:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:38
BLANK TEST	0.00	10:39

