

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 10/12/2020

OPERATOR NAME:
 JACQUELINE J BLUMFIELD
 Permit #: 38047
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:24
INTERNAL STANDARD	VERIFIED	10:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:25
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:26
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:27
BLANK TEST	0.00	10:28

