

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 10/19/2020

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:46
INTERNAL STANDARD	VERIFIED	09:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:47
BLANK TEST	0.00	09:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:48
BLANK TEST	0.00	09:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:49
BLANK TEST	0.00	09:50

