

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 11/02/2020

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:03
INTERNAL STANDARD	VERIFIED	11:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:03
BLANK TEST	0.00	11:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:04
BLANK TEST	0.00	11:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:06
BLANK TEST	0.00	11:06

