

SOLUTION CHANGE

SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106

Date: 11/17/2020

OPERATOR NAME:  
 MATTHEW J BOWE  
 Permit #: 40643  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:29
INTERNAL STANDARD	VERIFIED	08:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:32

*Handwritten signature*  
*MC*

