

SOLUTION CHANGE

SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106

Date: 11/23/2020

OPERATOR NAME:  
 KHALID PARWANTA  
 Permit #: 31498  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:23
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:24
BLANK TEST	0.00	08:25

