

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 12/07/2020

OPERATOR NAME:
 KHALID PARWANTA
 Permat #: 31498
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:23
INTERNAL STANDARD	VERIFIED	12:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:26
BLANK TEST	0.00	12:27

