

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 12/21/2020

OPERATOR NAME:
 PHILIP E MISSERE
 Permit #: 41131
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:20
INTERNAL STANDARD	VERIFIED	08:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:22
BLANK TEST	0.00	08:23

