

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 01/04/2021

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:21 |
| INTERNAL STANDARD | VERIFIED | 08:21 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.10 | 08:21 |
| BLANK TEST | 0.00 | 08:22 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:22 |
| BLANK TEST | 0.00 | 08:23 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:24 |
| BLANK TEST | 0.00 | 08:25 |

*P.O.
M J Bowe*

