

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 01/11/2021

OPERATOR NAME:  
ANDREW S COLE-HATCHARD  
Permit #: 47565  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:02
INTERNAL STANDARD	VERIFIED	10:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:02
BLANK TEST	0.00	10:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:03
BLANK TEST	0.00	10:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:05
BLANK TEST	0.00	10:05

