

SOLUTION CHANGE

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 02/01/2021

OPERATOR NAME:  
RANDY DIPAOLO  
Permit #: 45742  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31

