

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 02/08/2021

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:55
INTERNAL STANDARD	VERIFIED	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:55
BLANK TEST	0.00	09:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:59

[Handwritten Signature]
 #620

