

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 02/15/2021

OPERATOR NAME:  
MATTHEW J BOWE  
Permit #: 40643  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:20
INTERNAL STANDARD	VERIFIED	09:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:23
BLANK TEST	0.00	09:24


