

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 03/01/2021

OPERATOR NAME:
 MATTHEW MCALLISTER
 Permit #: 47389
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:41
INTERNAL STANDARD	VERIFIED	14:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:41
BLANK TEST	0.00	14:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:42
BLANK TEST	0.00	14:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:43
BLANK TEST	0.00	14:44

