

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106

Date: 03/08/2021

OPERATOR NAME:  
 KHALID PARWANTA  
 Permit #: 31498  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:33
INTERNAL STANDARD	VERIFIED	13:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:35
BLANK TEST	0.00	13:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:36
BLANK TEST	0.00	13:37

