

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 03/15/2021

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:55
INTERNAL STANDARD	VERIFIED	11:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:55
BLANK TEST	0.00	11:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:56
BLANK TEST	0.00	11:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:57
BLANK TEST	0.00	11:58

*P.O.
 M J
 H 620*

