

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 03/22/2021

OPERATOR NAME:  
KHALID PARWANTA  
Permit #: 31498  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:49
INTERNAL STANDARD	VERIFIED	08:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:51
BLANK TEST	0.00	08:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:52
BLANK TEST	0.00	08:53

