

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 04/05/2021

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:29
INTERNAL STANDARD	VERIFIED	19:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:30
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:31
BLANK TEST	0.00	19:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:32
BLANK TEST	0.00	19:33

