

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 04/12/2021

OPERATOR NAME:
 ANDREW S COLE-HATCHARD
 Permit #: 47565
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:11 |
| INTERNAL STANDARD | VERIFIED | 08:11 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:12 |
| BLANK TEST | 0.00 | 08:13 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:13 |
| BLANK TEST | 0.00 | 08:14 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:14 |
| BLANK TEST | 0.00 | 08:15 |

