

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 04/19/2021

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:29
INTERNAL STANDARD	VERIFIED	14:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:29
BLANK TEST	0.00	14:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:31
BLANK TEST	0.00	14:32

