

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 05/04/2021

OPERATOR NAME:
 PHILIP E MISSERE
 Permit #: 41131
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:46
INTERNAL STANDARD	VERIFIED	08:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50

