

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 05/17/2021

OPERATOR NAME:  
ANDREW S COLE-HATCHARD  
Permit #: 47565  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:38
INTERNAL STANDARD	VERIFIED	20:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:39
BLANK TEST	0.00	20:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:41
BLANK TEST	0.00	20:42

